

# Repeatability of Corneal Aberrations and Ocular Biometry Measurements Using the Pentacam® AXL: A Cross-sectional Study

MEZNAH S ALMUTAIRI<sup>1</sup>, BASAL H ALTOAIMI<sup>2</sup>, RANIA FAHMY<sup>3</sup>, SHATHA I ALGOWAIFLY<sup>4</sup>, MAY S ALENAZI<sup>5</sup>

## ABSTRACT

**Introduction:** Accurate assessment of ocular biometry and corneal aberrations for diagnosis and therapy is crucial in many clinical settings. Devices to improve the measurements of these parameters are constantly being developed.

**Aim:** To evaluate the repeatability of ocular biometry and corneal aberrations obtained by the Pentacam® AXL in normal eyes.

**Materials and Methods:** This cross-sectional study was conducted in the Optometry Department of King Saud University, Riyadh, Saudi Arabia from September 2021 to February 2022. The study included a total of 120 eyes from 60 normally sighted adult participants whose ocular biometry and corneal aberrations were measured three times in a single session using the Pentacam® AXL. Biometric parameters include Axial Length (AL), mean K-reading (Km), Anterior Chamber Depth (ACD), Central Corneal Thickness (CCT), and Horizontal White-to-White Distance (HWTW). The aberrometry parameters include

the most significant corneal higher-order aberrations (coma, trefoil, spherical aberration), and the Root Mean Square (RMS) of the higher-order aberrations. Repeatability was assessed using within-subject Standard Deviation (SD<sup>w</sup>), the repeatability limit (r), and the Intraclass Correlation Coefficient (ICC).

**Results:** The current study included 120 eyes of 60 subjects with a mean age of 21±2.08 years (range: 18 to 28 years). The repeatability was good for all recorded biometric measurements, with the most repeatable being AL (r=0.028, SD=0.01). All values of ICC for biometric measurements revealed excellent repeatability, being over 0.994. In addition, coma and spherical aberrations had excellent repeatability with ICC ≥0.935, and RMS, trefoil, secondary astigmatism, and quadrafoil aberrations had good repeatability with ICC 0.938, 0.823, 0.898, and 0.827, respectively.

**Conclusion:** The Pentacam® AXL demonstrated good repeatability for ocular biometry and corneal aberrations in healthy eyes.

**Keywords:** Axial length, Non contact biometry, Ocular aberrations, Partial coherence interferometry

## INTRODUCTION

Accurate and reliable measurements of ocular biometry parameters (such as corneal curvature, ACD, and AL) are essential in several clinical procedures and critical to surgical success and patient satisfaction. These parameters are essential for the Intraocular Lens (IOL) power calculation formula.

Ultrasound measurement has historically been the standard for calculating an IOL power [1,2]. However, technologies are constantly being developed in clinical practice to improve the precision measurement of these parameters for performing cataract and refractive surgery. These instruments have different measurement-based systems—for example, partial coherence interferometry (e.g., IOLMaster), optical low coherence reflectometry (e.g., Lenstar 900), and swept-source optical coherence tomography (e.g., IOLmaster 700) [3,4]. The Pentacam® AXL (Oculus, Wetzlar, Germany) has integrated non contact biometry using partial coherence interferometry and has an advantage over the widely used IOL Master by considering the posterior corneal curvature and reporting the total corneal refractive power [5-7].

In normal eyes, the cornea contributes 90% of total aberrations [8]. The measurement of corneal aberrations has long been of interest [9]. The Pentacam® AXL calculates the corneal aberrations based on corneal elevation data, fitting Zernike polynomials to the measured height data. A more accurate fitting will be achieved as more polynomial terms are used to define the aberration profile. Although several studies demonstrate the repeatability of this instrument in the measurement of ocular biometry [7,10], to author's knowledge, none have evaluated its capability to provide repeatable measurements of corneal aberrations.

Moreover, most studies on corneal aberrations focus only on the anterior corneal surface and ignore the posterior corneal surface

[11-14]. However, the posterior corneal surface must be considered to describe corneal aberrations accurately [15]. Due to the lack of comprehensive assessment published to date, the present study aimed to evaluate the repeatability of corneal aberrations and ocular biometry measurements obtained by the Pentacam® AXL in normal eyes with a transparent lens.

## MATERIALS AND METHODS

This cross-sectional study was conducted in the Optometry Department of King Saud University, Riyadh, Saudi Arabia from September 2021 to February 2022. The local Ethical Committee of King Saud University previously approved the study (Ref. No. 21/0460/IRB), and it adheres to the Declaration of Helsinki. Written informed consent was obtained from all subjects after they had been informed of the nature of the study.

**Inclusion criteria:** The study included normally sighted participants who had not used contact lenses in the two weeks before the examination.

**Exclusion criteria:** The study excluded subjects with a history of ocular surgery, trauma, or ocular pathology (example- no dry eye, corneal pathology, cataracts, or retinal disease).

## Study Procedure

For present study, convenience sampling was used. The present study included 120 eyes of 60 normally sighted participants, all of whom underwent a standardised ophthalmic examination. Authors used the Pentacam® AXL (Oculus, Wetzlar, Germany) with integrated non contact biometry using partial coherence interferometry technology for AL measurements. In combination with a Scheimpflug rotating camera, this device provides analysis of the anterior segment parameters and calculates the IOL power

required in cataract and refractive surgery. It uses a 475 nm blue Light-emitting Diode (LED) as a light source. The images of the cornea were captured with a 1.45-megapixel camera that records 138,000 data points within two seconds. In addition, Keratometry was calculated using a reference surface [7]. For the current study, authors used software version 1.22r05.

Authors calibrated the Pentacam® AXL as recommended by the manufacturer before use; all measurements were obtained after this calibration. A trained examiner took three measurements from both eyes of each subject under standardised conditions to minimise bias. These conditions included taking the measurements in the same dim room, a minute-long interval between each measurement in which patients were asked to stand up (the position of the joystick was changed), and having patients blink normally to avoid tear film disturbance before each measurement. The measures were considered acceptable according to the device manufacturer's quality criteria (they were marked as "OK" by an automatic quality check).

The ocular biometry and monochromatic corneal aberrations were measured without pupil dilation. A total of 25 images per scan were acquired to produce high-resolution corneal measurements. Two main groups of parameters, biometric and aberrometry, were measured. The biometric parameters included ACD, AL, CCT, HWTW, and Km. The aberrometry parameters included the most significant corneal higher-order aberrations (coma, trefoil, spherical aberration) [12] and the RMS of the higher-order aberrations. The higher-order aberrations were reported using the convention of the standards for reporting the optical aberrations [16].

## STATISTICAL ANALYSIS

All data were exported into an Excel file (Microsoft Inc, Redmond, WA, USA) and transferred to Statistical Packages for Social Sciences (SPSS) 28.0 (SPSS Inc, Chicago, IL, USA) for statistical analysis. After the descriptive analysis was performed, the repeatability of the Pentacam® AXL with three consecutive measurements was evaluated using the following:

- **Mean of within-subjects Standard Deviation (SDw) [17]:** The SD of the three repeated measures of each parameter was calculated for each eye, and then the mean of these deviations was obtained to generate the SD. A lower SDw indicates higher repeatability.
- **The repeatability limit (r):** This is reported as  $(r = \sqrt{2} \times 1.96 \text{ SD}^w)$ , considering the 95% confidence interval [17]. A lower value indicates a more repeatable the parameter.

- **Intraclass Correlation Coefficient (ICC):** This is the ratio of the between-subject variance to the pooled within-subject variance and the between-subject variance. It was automatically calculated using SPSS software with the two-way mixed effects Model (absolute agreement definition). ICC ranges from 0 to 1 and is commonly classified as acceptable, good, and excellent with greater than 0.7, greater than 0.8, and greater than 0.9, respectively [7,18].
- **One-sample t-tests:** These were performed after calculating the differences between each eye's first and second, first and third, and second and third measurements. A p-value less than 0.05 was considered statistically significant.

## RESULTS

The current study included 120 eyes of 60 subjects with a mean age of  $21 \pm 2.08$  years (range: 18 to 28 years). Their visual acuity was equal to or better than 20/20, with a mean spherical equivalent of  $0.54 \pm 0.43$ D for the right eye and  $0.65 \pm 0.56$ D for the left eye. [Table/Fig-1] describes the mean  $\pm$ SD values of all biometric parameters and corneal monochromatic aberrations for the three measurements and their comparisons.

**Biometric parameters:** The mean AL was  $23.63 \pm 1.12$  mm; the three measurements gave the same value, with a mean difference of  $-0.001 \pm 0.03$  mm between them. This was not statistically significant ( $p=0.776$ ; measurement 1 vs. 2;  $p=0.706$ ; measurement 1 vs. 3;  $p=0.319$ ; measurement 2 vs. 3). The mean for the K reading was  $43.07 \pm 1.41$ , with the differences between measurements being  $-0.0004 \pm 0.097$  D; these differences were not statistically significant ( $p=1$ ; measurement 1 vs. 2,  $p=0.869$ ; measurement 1 vs. 3  $p=0.318$ ; measurement 2 vs. 3). There was no statistical difference between mean ACD, CCT, and HWTW ( $p>0.05$ , [Table/Fig-2]).

**Corneal aberrations:** The mean of total aberrations (RMS) was  $1.47 \pm 0.41$   $\mu$ m. Lower-order aberrations had the highest RMS of  $1.41 \pm 0.41$   $\mu$ m, followed by higher-order aberrations ( $0.41 \pm 0.11$   $\mu$ m) and spherical aberrations ( $0.18 \pm 0.08$   $\mu$ m). There was no statistical difference between measurements in all values [Table/Fig-2].

**Repeatability:** The results of the SDw, the repeatability limit, and the intraclass correlation coefficient are reported in [Table/Fig-2]. The repeatability was good for all recorded measurements, with the most repeatable being AL ( $r=0.028$ ,  $\text{SDw}=0.01$ ). All ICC values for AL, Kmean, ACD, CCT, and HWTW measurements revealed excellent repeatability with an ICC over 0.994. Additionally, there was excellent repeatability of RMS, coma, and spherical aberrations, with  $\text{ICC} \geq$

Parameters	First measurement (m1)	Second measurement (m2)	Third measurement (m3)	Mean
<b>Ocular biometry</b>				
AL (mm)	23.63 $\pm$ 1.13	23.63 $\pm$ 1.12	23.63 $\pm$ 1.12	23.63 $\pm$ 1.12
Kmean (D)	43.07 $\pm$ 1.4	43.07 $\pm$ 1.41	43.06 $\pm$ 1.41	43.07 $\pm$ 1.41
ACD (mm)	3.49 $\pm$ 0.31	3.49 $\pm$ 0.32	3.49 $\pm$ 0.31	3.49 $\pm$ 0.31
CCT ( $\mu$ m)	554.24 $\pm$ 27.16	554.36 $\pm$ 27.21	554.67 $\pm$ 27.39	554.42 $\pm$ 27.25
HWTW (mm)	11.87 $\pm$ 0.39	11.86 $\pm$ 0.38	11.86 $\pm$ 0.39	11.86 $\pm$ 0.39
<b>Corneal aberrations</b>				
RMS ( $\mu$ m)	1.47 $\pm$ 0.40	1.47 $\pm$ 0.44	1.48 $\pm$ 0.39	1.47 $\pm$ 0.41
Higher-order aberrations RMS ( $\mu$ m)	0.41 $\pm$ 0.11	0.41 $\pm$ 0.10	0.41 $\pm$ 0.11	0.41 $\pm$ 0.11
Lower-order aberrations RMS ( $\mu$ m)	1.40 $\pm$ 0.40	1.41 $\pm$ 0.44	1.42 $\pm$ 0.39	1.41 $\pm$ 0.41
Coma aberrations ( $\mu$ m)	0.042 $\pm$ 0.19	0.04 $\pm$ 0.19	0.038 $\pm$ 0.19	0.04 $\pm$ 0.19
Trefoil aberrations ( $\mu$ m)	-0.057 $\pm$ 0.1	-0.05 $\pm$ 0.1	-0.05 $\pm$ 0.1	-0.05 $\pm$ 0.1
Spherical aberrations ( $\mu$ m)	0.18 $\pm$ 0.08	0.18 $\pm$ 0.08	0.18 $\pm$ 0.08	0.18 $\pm$ 0.08
Secondary astigmatism aberrations ( $\mu$ m)	-0.023 $\pm$ 0.05	-0.023 $\pm$ 0.05	-0.022 $\pm$ 0.05	-0.023 $\pm$ 0.05
Quadrafoil aberrations ( $\mu$ m)	-0.024 $\pm$ 0.07	-0.024 $\pm$ 0.06	-0.021 $\pm$ 0.06	-0.023 $\pm$ 0.07

**[Table/Fig-1]:** Mean  $\pm$ Standard Deviation (SD) values of all biometric parameters and corneal monochromatic aberrations.

AL: Axial length; Km: Mean keratometry; ACD: Anterior chamber depth; CCT: Central corneal thickness; HWTW: Horizontal white-to-white distance; RMS: Root mean square

Parameters	SDw	r	ICC	p-value (m1-m2)	p-value (m1-m3)	p-value (m2-m3)
<b>Ocular biometry</b>						
AL (mm)	0.01	0.028	1	0.776	0.706	0.319
Kmean (D)	0.05	0.149	0.999	1	0.869	0.318
ACD (mm)	0.02	0.047	0.999	0.976	0.927	0.328
CCT ( $\mu\text{m}$ )	1.8	5.00	0.998	0.634	0.118	0.314
HWTW (mm)	0.04	0.105	0.994	0.361	0.243	0.320
<b>Corneal aberrations</b>						
RMS ( $\mu\text{m}$ )	0.115	0.319	0.938	0.891	0.610	0.503
Higher-order aberrations RMS ( $\mu\text{m}$ )	0.034	0.094	0.936	0.890	0.886	0.484
Lower-order aberrations RMS ( $\mu\text{m}$ )	0.115	0.321	0.935	0.876	0.565	0.525
Coma aberrations ( $\mu\text{m}$ )	0.026	0.073	0.984	0.791	0.448	0.534
Trefoil aberrations ( $\mu\text{m}$ )	0.052	0.144	0.823	0.478	0.455	0.449
Spherical aberrations ( $\mu\text{m}$ )	0.020	0.056	0.960	0.433	0.336	0.351
Secondary astigmatism aberrations ( $\mu\text{m}$ )	0.020	0.056	0.898	0.532	0.312	0.181
Quadrafoil aberrations ( $\mu\text{m}$ )	0.034	0.093	0.827	0.572	0.509	0.318

**[Table/Fig-2]:** Repeatability analysis of the three consecutive measurements.

AL: axial length; K<sub>mean</sub>: mean keratometry; ACD: Anterior chamber depth; HWTW: Horizontal white-to-white distance; SD<sup>w</sup>: within-subject standard deviation; r: the repeatability limit ( $r = \sqrt{2} \times 1.96 \text{ SD}^w$ ); ICC: Intraclass correlation coefficient. \* p-values were calculated by one sample t test

0.935, and good repeatability of trefoil, secondary astigmatism, and quadrafoil aberrations with an ICC of 0.823, 0.898, and 0.827, respectively.

According to the repeatability limits obtained when measuring each biometric parameter [Table/Fig-2], the repeatability was very good for all recorded measurements. The least repeatable parameter was CCT (5.00), while the most repeatable was AL (0.028). Additionally, based on the repeatability limits obtained when measuring corneal aberrations [Table/Fig-2], the repeatability was very good for all recorded aberration measurements. The least repeatable parameter was lower-order aberrations RMS (0.321), and the most repeatable were spherical aberrations and secondary astigmatism aberrations (0.056). These results indicate that the Pentacam® AXL measures these parameters with high repeatability.

## DISCUSSION

The present study assessed the repeatability of ocular biometry and corneal aberrations measurements obtained by the Pentacam® AXL in normal eyes. Three repeated measures were collected from both eyes of each subject. The differences between the measurements were not statistically significant for any of the parameters studied.

The mean difference between AL measurements was  $0.001 \pm 0.03$  mm, which theoretically could have an insignificant influence on IOL power measurement and subsequent implantation. The Pentacam® AXL showed excellent repeatability, established by ICC values greater than 0.994 for all biometric parameters. These results are comparable to those found by Srivannaboon S et al., who reported an ICC higher than 0.974 in all measurements [1]. Similar to the current study, the lowest ICC value was for the measure of the HWTW distance.

A variability in ACD measurement of 0.02 mm would yield a movement of less than 0.10 mm in the IOL position, which could amount to approximately 0.20 D. As proved by Olsen T and Hoffmann P an error less than  $\pm 0.50$  D is deemed optimal in phaco-refractive surgery [19]. In addition to IOL calculation, AL and ACD measurements may assist in assessing patients with narrow-angle glaucoma. Nongpiur ME et al., proved that the difference between healthy and angle-closure glaucoma patients was 0.30 mm, a much higher value than the variability presented by the Pentacam® AXL, which was 0.02 mm. The ACD measurement, ICC, and SDw values reported in present study were consistent with those revealed by Grulkowski I et al., employing long-range swept-source optical

coherence tomography using a Vertical-cavity Surface-emitting Laser (VCSEL), achieving an ICC of 0.994 [20,21].

Keratometry repeatability data (mean keratometry 0.05 D) were not statistically significant; it could clinically yield a variation of 0.25 D in the IOL power. Test-retest repeatability confirmed that corneal power measurements were repeatable and expected to be less than a quarter of a diopter in 95% of paired observations, which would have little impact clinically.

This value has a low clinical impact on calculating IOL power [22]. The ICC and SDw values were in agreement with Lo'pez de la Fuente C et al., who compared anterior segment measurements acquired from healthy subjects with three different devices, including the IOLMaster 500 [23].

The HWTW and ACD are two fundamental values for posterior chamber phakic IOLs (pIOL) calculation. In particular, the HWTW is utilised to compute the power of the pIOL to be implanted. This parameter is crucial because if a smaller pIOL than needed is implanted, it can be complicated by cataracts [24]; furthermore, if a larger pIOL is implanted, it can be complicated by angle-closure glaucoma [25]. Additionally, the difference between the HWTW and the power of the lens sets the vault [26]. The present measurements of white-to-white distance ( $11.86 \pm 0.39$  mm) were slightly higher than those recorded by Martin R et al., ( $11.47 \pm 0.36$  mm) in myopic patients with implanted pIOL; however, they used scanning-slit topography-based technology, which provides lower values than the Pentacam® AXL [27].

The CCT values ( $554.42 \pm 27.25$  microns) were similar to those estimated in the healthy population in Mexico by de la Parra P et al., ( $542.333 \pm 3.446$  microns) with Scheimpflug Sirius camera tomography [28].

Some studies have documented the repeatability of corneal aberrometric measurements [29-32], but none focused on internal aberrometry. The present study found no significant differences in higher-order aberrations, suggesting constant instrument repeatability over time. This was similar to the findings of Miranda MA et al., who interpreted aberration data over one week [30]. Similarly, they obtained repeatability values and found the highest variation in the total higher-order aberrations. Visser N et al., also revealed good repeatability values for two Hartmann-Shack aberrometers, the IRIX-3 (Imagine Eyes, Paris, France), and the Keratron Onda (Optikon, Rome, Italy), with mean SD values below  $0.1 \mu\text{m}$  for both aberrometers for all higher-order aberrations and slightly larger values (0.1 to  $0.2 \mu\text{m}$ ) for second-



order aberrations [33]. In contrast, the SD was 0.034  $\mu\text{m}$  in the current study.

The ICC values acquired in present study were homogeneous or slightly higher than those established by Piñero DP et al., who recorded ICC values above 0.75 for most aberrations analysed from three repeated consecutive automatic measurements recorded by the same examiner [34]. No pupillary dilation was performed for the measurements, and a 4-mm pupil was used for computation. Variability was noticed in repeated measurements of RMS values presented by primary coma, secondary astigmatism, and tetrafoil. These parameters had lower ICC values in the current study's repeatability analysis.

### Limitation(s)

The present study had some limitations. Firstly, a power analysis was not performed to calculate the sample size, which could have affected the results. Additionally, the number of eyes examined and the inclusion of both eyes would have impacted the results due to their correlation. However, the main aim of the current study was to determine the repeatability of the Pentacam® AXL, which would be influenced by this correlation. Moreover, single-eye analysis demonstrated comparable results, and the one-eye process has the significant disadvantage of a loss of information; hence, many authors do not find it ideal [35,36]. It is recommended that repeated measurements of internal aberrations be done when the device is used clinically or for research to prevent variability in the measurement of some aberrometric errors.

### CONCLUSION(S)

In conclusion, the present study confirmed that the Pentacam® AXL has high repeatability when used to measure biometric parameters and corneal aberrations in healthy eyes. The Pentacam® AXL seems suitable for use as an all-inclusive optical biometer and corneal tomographer, as it combines partial coherence interferometry for obtaining accurate AL measurements with Scheimpflug technology to obtain ACD and K measurements.

### Acknowledgement

The authors extend their appreciation to the Deanship of Scientific Research, College of Applied Medical Sciences Research Center at King Saud University.

### REFERENCES

- [1] Srivannaboon S, Chirapapaisan C, Chonpimai P, Koodkaew S. Comparison of ocular biometry and intraocular lens power using a new biometer and a standard biometer. *J Cataract Refract Surg.* 2014;40(5):709-15. Available from: <https://doi.org/10.1016/j.jcrs.2013.09.020>.
- [2] Kaswin G, Rousseau A, Mgarrech M, Barreau E, Labetoulle M. Biometry and intraocular lens power calculation results with a new optical biometry device: Comparison with the gold standard. *J Cataract Refract Surg.* 2014;40(4):593-600. Available from: <https://doi.org/10.1016/j.jcrs.2013.09.015>.
- [3] Ayala M, Strandås R. Accuracy of optical coherence tomography (OCT) in pachymetry for glaucoma patients. *BMC Ophthalmol.* 2015;15:124. Available from: <https://doi.org/10.1186/s12886-015-0116-x>.
- [4] Shammas HJ, Ortiz S, Shammas MC, Kim SH, Chong C. Biometry measurements using a new large-coherence-length swept-source optical coherence tomographer. *J Cataract Refract Surg.* 2016;42(1):50-61. Available from: <https://doi.org/10.1016/j.jcrs.2015.07.042>.
- [5] Rydstrom E, Westin O, Koskela T, Behndig A. Posterior corneal astigmatism in refractive lens exchange surgery. *Acta Ophthalmol.* 2016;94(3):295-300. Available from: <https://doi.org/10.1111/aos.12965>.
- [6] Koch DD, Ali SF, Weikert MP, Shirayama M, Jenkins R, Wang L. Contribution of posterior corneal astigmatism to total corneal astigmatism. *J Cataract Refract Surg.* 2012;38(12):2080-87.
- [7] Sel S, Stange J, Kaiser D, Kiraly L. Repeatability and agreement of Scheimpflug-based and swept-source optical biometry measurements. *Cont Lens Anterior Eye.* 2017;40(5):318-22. Available from: <https://doi.org/10.1016/j.clae.2017.03.007>.
- [8] Karimian F, Feizi S, Doozande A. Higher-order aberrations in myopic eyes. *J Ophthalmic Vis Res.* 2010;5(1):03-09.
- [9] Atchison DA. Recent advances in representation of monochromatic aberrations of human eyes. *Clin Exp Optom.* 2004;87(3):138-48. Available from: <https://doi.org/10.1111/j.1444-0938.2004.tb03166.x>.
- [10] Shajari M, Cremonese C, Petermann K, Singh P, Müller M, Kohlen T. Comparison of axial length, corneal curvature, and anterior chamber depth measurements of 2 recently introduced devices to a known biometer. *Am J Ophthalmol.* 2017;178:58-64. Available from: <https://doi.org/10.1016/j.ajo.2017.02.027>.
- [11] Wang L, Dai E, Koch D, Nathoo A. Optical aberrations of the human anterior cornea. *J Cataract Refract Surg.* 2003; 29: 1514-1521. Available from: [https://doi.org/10.1016/S0886-3350\(03\)00467-X](https://doi.org/10.1016/S0886-3350(03)00467-X)
- [12] Artal P, Berrio E, Guirao A, Piers P. Contribution of the cornea and internal surfaces to the change of ocular aberrations with age. *J Opt Soc Am A Opt Image Sci Vis.* 2002;19(1):137-43. Available from: <https://doi.org/10.1364/josaa.19.000137>.
- [13] Artal P, Guirao A, Berrio E, Williams DR. Compensation of corneal aberrations by the internal optics in the human eye. *J Vis.* 2001;1(1):01-08. Available from: <https://doi.org/10.1167/1.1.1>.
- [14] Atchison DA. Anterior corneal and internal contributions to peripheral aberrations of human eyes. *J Opt Soc Am A Opt Image Sci Vis.* 2004;21(3):355-59. Available from: <https://doi.org/10.1364/josaa.21.000355>.
- [15] Marcos S. Aberrations and visual performance following standard laser vision correction. *J Refract Surg.* 2001;17(5):S596-601. Available from: <https://doi.org/10.3928/1081-597x-20010901-19>.
- [16] Thibos LN, Applegate RA, Schwiegerling JT, Webb R; VSIA Standards Taskforce Members. Vision science and its applications. Standards for reporting the optical aberrations of eyes. *J Refract Surg.* 2002;18(5):S652-60. Available from: <https://doi.org/10.3928/1081-597x-20020901-30>.
- [17] McAlinden C, Khadka J, Pesudovs K. Statistical methods for conducting agreement (comparison of clinical tests) and precision (repeatability or reproducibility) studies in optometry and ophthalmology. *Ophthalmic Physiol Opt.* 2011;31(4):330-38. Available from: <https://doi.org/10.1111/j.1475-1313.2011.00851.x>.
- [18] Savini G, Schiano-Lomoriello D, Hoffer KJ. Repeatability of automatic measurements by a new anterior segment optical coherence tomographer combined with Placido topography and agreement with 2 Scheimpflug cameras. *J Cataract Refract Surg.* 2018;44(4):471-78. Available from: <https://doi.org/10.1016/j.jcrs.2018.02.015>.
- [19] Olsen T, Hoffmann P. C constant: new concept for ray tracing-assisted intraocular lens power calculation. *J Cataract Refract Surg.* 2014;40(5):764-73. Doi: 10.1016/j.jcrs.2013.10.037.
- [20] Nupur ME, He M, Amerasinghe N, Friedman DS, Tay WT, Baskaran M, et al. Lens vault, thickness, and position in Chinese subjects with angle closure. *Ophthalmology.* 2011;118(3):474-79. Available from: <https://doi.org/10.1016/J.OPHTHA.2010.07.025>.
- [21] Grulkowski I, Liu JJ, Zhang JY, Potsaid B, Jayaraman V, Cable AE, et al. Reproducibility of a long-range swept-source optical coherence tomography ocular biometry system and comparison with clinical biometers. *Ophthalmology.* 2013;120(11):2184-90. Available from: <https://doi.org/10.1016/J.OPHTHA.2013.04.007>.
- [22] Savini G, Barboni P, Carbonelli M, Hoffer KJ. Repeatability of automatic measurements by a new Scheimpflug camera combined with Placido topography. *J Cataract Refract Surg.* 2011;37(10):1809-16.
- [23] Lopez De La Fuente C, Sanchez-Cano A, Segura F, Pinilla I. Comparison of anterior segment measurements obtained by three different devices in healthy eyes. *Biomed Res Int.* 2014;2014:498080. Available from: <https://doi.org/10.1155/2014/498080>.
- [24] Gonvers M, Bornet C, Othenin-Girard P. Implantable contact lens for moderate to high myopia: Relationship of vaulting to cataract formation. *J Cataract Refract Surg.* 2003;29(5):918-24. Available from: [https://doi.org/10.1016/S0886-3350\(03\)00065-8](https://doi.org/10.1016/S0886-3350(03)00065-8).
- [25] Zaldivar R, Davidoff JM, Oscherow S. Posterior chamber phakic intraocular lens for myopia of -8 to -19 diopters. *Journal of Refractive Surgery.* 1998;14(3):294-305. Available from: <https://doi.org/10.3928/1081-597X-19980501-13>.
- [26] Alfonso JF, Lisa C, Palacios A, Fernandes P, González-Méjome JM, Montés-Micó R. Objective vs subjective vault measurement after myopic implantable collamer lens implantation. *Am J Ophthalmol.* 2009;147(6):978-83.e1. Available from: <https://doi.org/10.1016/J.AJO.2009.01.006>.
- [27] Martin R, Ortiz S, Rio-Cristobal A. White-to-white corneal diameter differences in moderately and highly myopic eyes: Partial coherence interferometry versus scanning-slit topography. *J Cataract Refract Surg.* 2013;39(4):585-89. Available from: <https://doi.org/10.1016/J.JCRS.2012.11.021>.
- [28] De La Para-Colin P, Garza-León M, Barrientos-Gutierrez T. Repeatability and comparability of anterior segment biometry obtained by the Sirius and the Pentacam analyzers. *Int Ophthalmol.* 2014;34(1):27-33. Available from: <https://doi.org/10.1007/S10792-013-9780-0>.
- [29] Holzer MP, Sassenroth M, Auffarth GU. Reliability of corneal and total wavefront aberration measurements with the SCHWIND corneal and ocular wavefront analyzers. *J Refract Surg.* 2006;22(9):917-20. Available from: <https://doi.org/10.3928/1081-597X-20061101-16>.
- [30] Miranda MA, O'Donnell C, Radhakrishnan H. Repeatability of corneal and ocular aberration measurements and changes in aberrations over one week. *Clin Exp Optom.* 2009;92(3):253-66. Available from: <https://doi.org/10.1111/j.1444-0938.2009.00364.x>.
- [31] Wang L, Shirayama M, Koch DD. Repeatability of corneal power and wavefront aberration measurements with a dual-Scheimpflug Placido corneal topographer. *J Cataract Refract Surg.* 2010;36(3):425-30. Available from: <https://doi.org/10.1016/J.JCRS.2009.09.034>.

- [32] Gobbe M, Guillon M, Maissa C. Measurement repeatability of corneal aberrations. *J Refract Surg.* 2002;18(5):S567-71. Available from: <https://doi.org/10.3928/1081-597X-20020901-14>.
- [33] Visser N, Berendschot TTJM, Verbakel F, Tan AN, de Brabander J, Nuijts RMMA. Evaluation of the comparability and repeatability of four wavefront aberrometers. *Invest Ophthalmol Vis Sci.* 2011;52(3):1302-11. Available from: <https://doi.org/10.1167/IOVS.10-5841>.
- [34] Piñero DP, Juan JT, Alió JL. Intrasubject repeatability of internal aberrometry obtained with a new integrated aberrometer. *J Refract Surg.* 2011;27(7):509-17. Available from: <https://doi.org/10.3928/1081597X-20101214-01>.
- [35] Murdoch IE, Morris SS, Cousens SN. People and eyes: Statistical approaches in ophthalmology. *Br J Ophthalmol.* 1998;82(8):971-73. Available from: <https://doi.org/10.1136/BJO.82.8.971>.
- [36] Karakosta A, Vassilaki M, Plainis S, Elfadl NH, Tsilimbaris M, Moschandreas J. Choice of analytic approach for eye-specific outcomes: One eye or two? *Am J Ophthalmol.* 2012;153(3):571-79.e1. Available from: <https://doi.org/10.1016/j.ajo.2011.08.032>.

**PARTICULARS OF CONTRIBUTORS:**

1. Assistant Professor, Department of Optometry, King Saud University, Riyadh, Saudi Arabia.
2. Assistant Professor, Department of Optometry, King Saud University, Riyadh, Saudi Arabia.
3. Associate Professor, Department of Optometry, King Saud University, Riyadh, Saudi Arabia.
4. Lecturer, Department of Optometry, King Saud University, Riyadh, Saudi Arabia.
5. O.D, Department of Optometry, King Saud University, Riyadh, Saudi Arabia.

**NAME, ADDRESS, E-MAIL ID OF THE CORRESPONDING AUTHOR:**

Meznah S Almutairi,  
College of Applied Medical Sciences, King Saud University, Riyadh-11433, Saudi Arabia.  
E-mail: mzalmutairi@ksu.edu.sa

**PLAGIARISM CHECKING METHODS:** [\[Jain H et al.\]](#)

- Plagiarism X-checker: Oct 17, 2023
- Manual Googling: Mar 11, 2024
- iThenticate Software: Mar 13, 2024 (11%)

**ETYMOLOGY:** Author Origin**EMENDATIONS:** 7**AUTHOR DECLARATION:**

- Financial or Other Competing Interests: None
- Was Ethics Committee Approval obtained for this study? Yes
- Was informed consent obtained from the subjects involved in the study? Yes
- For any images presented appropriate consent has been obtained from the subjects. NA

Date of Submission: **Oct 16, 2023**  
Date of Peer Review: **Nov 30, 2023**  
Date of Acceptance: **Mar 16, 2024**  
Date of Publishing: **May 01, 2024**